



Paper-Based Electrochemical Biosensors for Voltammetric Detection of miRNA Biomarkers Using Reduced Graphene Oxide or MoS₂ Nanosheets Decorated with Gold Nanoparticle Electrodes

Torul, H., Yarali, E., Eksin, E., Ganguly, A., Benson, J., Tamer, U., Papakonstantinou, P., & Erdem, A. (2021). Paper-Based Electrochemical Biosensors for Voltammetric Detection of miRNA Biomarkers Using Reduced Graphene Oxide or MoS₂ Nanosheets Decorated with Gold Nanoparticle Electrodes. *Biosensors*, 11(7), [236]. <https://doi.org/10.3390/bios11070236>, <https://doi.org/10.3390/bios11070236>

[Link to publication record in Ulster University Research Portal](#)

Published in:
Biosensors

Publication Status:
Published (in print/issue): 13/07/2021

DOI:
[10.3390/bios11070236](https://doi.org/10.3390/bios11070236)
[10.3390/bios11070236](https://doi.org/10.3390/bios11070236)

Document Version
Publisher's PDF, also known as Version of record

General rights

Copyright for the publications made accessible via Ulster University's Research Portal is retained by the author(s) and / or other copyright owners and it is a condition of accessing these publications that users recognise and abide by the legal requirements associated with these rights.

Take down policy

The Research Portal is Ulster University's institutional repository that provides access to Ulster's research outputs. Every effort has been made to ensure that content in the Research Portal does not infringe any person's rights, or applicable UK laws. If you discover content in the Research Portal that you believe breaches copyright or violates any law, please contact pure-support@ulster.ac.uk.

Article

Paper-Based Electrochemical Biosensors for Voltammetric Detection of miRNA Biomarkers Using Reduced Graphene Oxide or MoS₂ Nanosheets Decorated with Gold Nanoparticle Electrodes

Hilal Torul ^{1,†} , Ece Yarali ², Ece Eksin ^{2,†} , Abhijit Ganguly ³, John Benson ⁴, Ugur Tamer ¹, Pagona Papakonstantinou ^{3,*}  and Arzum Erdem ^{2,*} 

¹ Department of Analytical Chemistry, Faculty of Pharmacy, Gazi University, Ankara 06330, Turkey; hilaltorul@gazi.edu.tr (H.T.); utamer@gazi.edu.tr (U.T.)

² Department of Analytical Chemistry, Faculty of Pharmacy, Ege University, Bornova 35100, Turkey; eceyarali@hotmail.com (E.Y.); eceksin@hotmail.com (E.E.)

³ School of Engineering, Engineering Research Institute, Ulster University, Newtownabbey BT37 0QB, UK; a.ganguly@ulster.ac.uk

⁴ 2-DTech, Core Technology Facility, 46 Grafton Street, Manchester M13 9NT, UK; john.benson@2-dtech.com

* Correspondence: p.papakonstantinou@ulster.ac.uk (P.P.); arzum.erdem@ege.edu.tr (A.E.)

† These authors contributed equally to this work.



Citation: Torul, H.; Yarali, E.; Eksin, E.; Ganguly, A.; Benson, J.; Tamer, U.; Papakonstantinou, P.; Erdem, A. Paper-Based Electrochemical Biosensors for Voltammetric Detection of miRNA Biomarkers Using Reduced Graphene Oxide or MoS₂ Nanosheets Decorated with Gold Nanoparticle Electrodes. *Biosensors* **2021**, *11*, 236. <https://doi.org/10.3390/bios11070236>

Received: 30 May 2021

Accepted: 5 July 2021

Published: 13 July 2021

Publisher's Note: MDPI stays neutral with regard to jurisdictional claims in published maps and institutional affiliations.



Copyright: © 2021 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).

Abstract: Paper-based biosensors are considered simple and cost-efficient sensing platforms for analytical tests and diagnostics. Here, a paper-based electrochemical biosensor was developed for the rapid and sensitive detection of microRNAs (miRNA-155 and miRNA-21) related to early diagnosis of lung cancer. Hydrophobic barriers to creating electrode areas were manufactured by wax printing, whereas a three-electrode system was fabricated by a simple stencil approach. A carbon-based working electrode was modified using either reduced graphene oxide or molybdenum disulfide nanosheets modified with gold nanoparticle (AuNPs/RGO, AuNPs/MoS₂) hybrid structures. The resulting paper-based biosensors offered sensitive detection of miRNA-155 and miRNA-21 by differential pulse voltammetry (DPV) in only 5.0 µL sample. The duration in our assay from the point of electrode modification to the final detection of miRNA was completed within only 35 min. The detection limits for miRNA-21 and miRNA-155 were found to be 12.0 and 25.7 nM for AuNPs/RGO and 51.6 and 59.6 nM for AuNPs/MoS₂ sensors in the case of perfectly matched probe-target hybrids. These biosensors were found to be selective enough to distinguish the target miRNA in the presence of single-base mismatch miRNA or noncomplementary miRNA sequences.

Keywords: paper-based biosensor; reduced graphene oxide; molybdenum disulfide nanosheets; microRNA; gold nanohybrids; differential pulse voltammetry

1. Introduction

The use of paper in chemical analysis started as early as the 1930s [1,2], and the first paper-based glucose sensor was fabricated in the 1950s [3]. However, paper-based sensors were identified as a distinctive category by Whiteside et al. in 2007 [4,5]. In the past decade, paper-based sensors have received increased interest because they are easy to use and disposable with low-cost fabrication [6–8]. They also provide benefits, such as short analysis time and usage of a small volume of sample [9]. Therefore, they are promising alternatives to traditional point-of-care devices. A typical paper-based electrochemical sensor consists of a paper as a substrate material, an electrode area, and two or three electrodes. To fabricate the electrode area, hydrophobic barriers are prepared using several techniques, such as chemical vapor-phase deposition, soft lithography, wax patterning, and inkjet printing [5]. Two- or three-electrode systems can be fabricated using various

techniques, such as photolithography [10], stencil printing [11], and inkjet printing [12]. Paper-based sensors find application in fields such as environmental analysis, biomedicine, food safety, chemical industry, and clinical analysis [13–17]. They can also be used as a type of point-of-care (POC) systems since they exhibit eco-friendly behavior [18–21].

Cancer is one of the most prevalent genetic diseases that leads to uncontrolled cell growth and deregulation of gene expression. miRNAs have been associated with cancer due to their wide impact on gene expression [22]. miRNAs have functions as oncogenes or tumor suppressors based on their inhibition of tumor-suppressive and oncogenic target mRNAs [23–25]. For example, miRNA-34, miRNA-126, miRNA-133, miRNA-143, and miRNA-145 are downregulated in many types of cancer. miRNA-15, miRNA-16, miRNA-21, miRNA-155, and miRNA-372 are found to be highly expressed in different types of tumors and promote oncogenesis [26]. Recent works have shown that miR-21 and miR-155 are the most significantly altered miRNAs in most cancer types [22,27,28]. Due to their critical roles in cancer and other diseases, miRNAs are considered crucial noninvasive biomarkers. Therefore, efforts have been made in recent years to develop innovative platforms for the efficient detection of microRNAs. Several methods, such as polymerase chain reaction (PCR)-based techniques, miRNA microarrays, sequencing, Northern blotting, mass spectrometry, optical and electrochemical methods have been examined for sensitive and selective miRNA detection [29–33].

Among them, electrochemical detection techniques have been considered as cost-effective and, at the same time, can provide high selectivity and sensitivity detection. Advances in electrochemical biomolecular detection are well discussed in a work reported by Sage et al. [34]. Paper-based sensors can be perfectly integrated with electrochemical techniques [35]. These combined systems enable the integration of different electrodes, such as ion-selective or microelectrodes [36,37]. They provide signal amplification while working in flow injection systems. Besides, paper-based analytical devices can also be coupled with several electrochemical methods, such as impedimetric detection, amperometry, cyclic voltammetry, coulometry, and potentiometric techniques [5].

In the present study, paper-based electrochemical biosensors based on gold nanoparticle hybrids were developed for selective and sensitive miRNA-155 and miRNA-21 detection. A carbon-based working electrode of a paper-based biosensor was modified using nanosheets of either reduced graphene oxide or molybdenum disulfide decorated with gold nanoparticles (denoted as AuNPs/RGO and AuNPs/MoS₂, respectively). The sequence of a complementary miRNA target was detected by a thiol-linked synthetic DNA probe, immobilized onto the working electrode by differential pulse voltammetry (DPV) in a redox [Fe(CN)₆]^{3−/4−} solution. According to the differentiation at the signal of redox probe measured in the absence/presence of miRNA hybridization on the paper-based biosensor, hybridization was detected.

2. Materials and Methods

2.1. Apparatus

Electrochemical measurements were carried out with an Autolab-302 PGSTAT and GPES 4.9.007 software package (Eco Chemie, Utrecht, The Netherlands). A Faraday cage (Eco Chemie, Utrecht, The Netherlands) was used to reduce background signal. Raw data were treated with a Savitzky–Golay filter (level 2) and moving average baseline correction (peak width, 0.03).

2.2. Chemicals


Carbon paste was purchased from Daejoo Electronic Materials Co., LTD. (Siheung-si, Korea). Ag/AgCl (9:1) ink was purchased from Henkel (Dusseldorf, Germany). Reduced graphene oxide (RGO) was produced by 2-DTech Ltd. (Manchester, UK) using a proprietary approach. Nitrocellulose (NC) membrane (Hi-Flow Plus HFC07504) was provided by Merck (Darmstadt, Germany). *N*-hydroxysuccinimide (NHS), chloroauric acid (HAuCl₄), and *N*-(3-dimethylaminopropyl)-*N*'-ethylcarbodiimide hydrochloride (EDC) were obtained

from Sigma-Aldrich (St. Louis, MI, USA). The EDC/NHS solution was prepared at 10.0 mM concentration for each component in pH 7.4 phosphate buffer. miRNAs and the base sequences of all oligonucleotides are given in the supporting information.

All other reagents were purchased from Sigma-Aldrich and Merck.

2.3. Generation and Modification of Paper Electrode

First, the paper electrode was developed as reported in our previous work [38]. It was constructed using a nitrocellulose membrane. After the construction of a pattern including a fluidic channel and electrode assembly area, a hydrophobic barrier was generated by utilizing a wax printer onto the NC membrane. Channels with a diameter of 2.0 mm and a length of 1.5 cm were constructed for capillary flow, and the length of the resulting channel was 0.6 cm. We placed three electrodes assembled in the working area designed with dimensions of $\sim 20 \text{ mm}^2$ and a 270 angle to obtain the maximum spread speed of the liquid. A pattern was designed onto a steel wafer of 0.1 mm thickness using a laser cutter. The resulting mask was placed on the NC membrane, and commercial carbon ink was used to create the working and counter electrodes. For the pseudo-reference electrode, an Ag/AgCl ink was used, and copper wires were used as conductive pads. The resulting electrode assembly was baked at 100 °C for 5 min. A schematic illustration of the electrode assembly is indicated in Scheme 1.



Scheme 1. Preparation of a paper electrode.

RGO powder was dispersed in ultrapure water at 1.0 mg/mL by a sonicator for 2 h. In order to obtain an RGO-modified paper electrode, 3.0 μL of 1.0 mg/mL RGO aqueous solution was applied on the surface of a working electrode three times. Between each drop of RGO aqueous solution, the electrode surface was dried under tungsten lamp for 5 min.

Chemical activation was carried out with EDC/NHS as a cross-linking agent. Each RGO-modified paper electrode surface was covered with 5.0 μL of EDC/NHS solution and interacted for 20 min to activate carboxyl groups on the surface of the RGO-modified paper electrode.

2.4. Preparation of a Molybdenum Disulfide Nanosheet (MoS_2)-Modified Paper Electrode

In this study, few-layer MoS_2 nanosheets were prepared by ionic-liquid-assisted grinding exfoliation, followed by sequential centrifugation steps, as described in our previous studies [39]. MoS_2 powder was dispersed in ultrapure water at 2.0 mg/mL by a

sonicator. The surface of the working electrode was covered by dropping 3.0 μL of MoS_2 aqueous solution. Then, the electrode surface was dried under tungsten lamp for 5 min.

2.5. Electrodeposition of Gold Nanoparticles (AuNPs) on an RGO- or MoS_2 -Modified Paper Electrode

After the preparation of an RGO- or MoS_2 -modified paper electrode, AuNPs were deposited onto the modified paper electrodes using the chronoamperometric technique in aqueous solution of a HAuCl_4 gold precursor by applying -0.3 V for 10 min.

2.6. miRNA Detection with AuNP/RGO- or AuNP/ MoS_2 -Modified Paper Electrodes

The surface of AuNP/RGO- or AuNP/ MoS_2 -modified paper electrodes was covered with 5.0 μL of thiol-linked Probe-1 or Probe-2. The DNA probe was covalently immobilized onto AuNPs. Then, a washing step was applied using PBS (pH 7.4) to prevent nonspecific binding. The hybridization of the probe and target microRNAs was achieved by dropping 5.0 μL of miRNA-155, or miRNA-21, on the surface of the electrodes. After the hybridization step, the prepared electrodes were washed with PBS in order to eliminate nonspecific adsorption.

2.7. Voltammetric Measurement

The experiments were carried out in 20.0 μL of a 1.0 mM $[\text{Fe}(\text{CN})_6]^{3-/4-}$ redox probe between -0.1 and $+0.3\text{ V}$ at a scan rate of 50 mV/s and a pulse amplitude of 50 mV by DPV.

CV measurements were performed by scanning between -1.0 and $+1.0\text{ V}$ at a scan rate of 50 mV/s in a redox probe solution of 50.0 mM $[\text{Fe}(\text{CN})_6]^{3-/4-}$ prepared in 0.1 M KCl.

3. Results and Discussion

3.1. Characterization Studies of the Paper Electrode Modified with Gold Nanoparticles/Reduced Graphene Oxide (AuNPs/RGO)

The characterization of the modified paper electrodes was achieved by Raman spectroscopy. A Raman microscope (DeltaNu Inc., Laramie, WY, USA) with a charge-coupled device detector, a laser source at 785 nm, and a motorized XYZ microscope stage specimen holder was utilized to characterize the working electrode surface. The measurements were achieved by using a 10X objective with a laser spot size of 7.5 μm . Raman signals were obtained with a laser power of 140 mW for an acquisition time of 20 s.

Raman spectra of the RGO-modified paper electrodes are shown in Figure 1A. The Raman peak of RGO at 1308 cm^{-1} was attributed to the D band correlated with the structural defects or disorders in the lattice structure. The band at 1590 cm^{-1} was related to the G band associated with the first-order scattering of the E_{2g} vibrational mode [40,41]. Gold interfacing on RGO enhanced the intensity of the D and G bands by 79.2% and 78.7%, respectively. The enhancement of the signals can be via the excitation of localized surface plasmons or the formation of charge-transfer complexes between RGO and AuNPs [42].

The morphological characterization of RGO- and AuNP/RGO-modified paper electrodes was realized using a Quanta 200 3D scanning electron microscope (SEM). As shown in Figure 1B(b), the resulting AuNPs were homogeneously dispersed onto the RGO surface. The size of the gold nanoparticle was found to be $229 \pm 53\text{ nm}$ and covered both RGO flakes and the working electrode area. These results demonstrate that AuNPs can be successfully electrodeposited onto agglomerates of RGO. A typical SEM image of RGO is shown in Figure 2, revealing a crumple-like morphology.

The electrochemical characterization of an unmodified paper electrode, RGO-modified paper electrode, and AuNP/RGO-modified paper electrode was performed by cyclic voltammetry (Figure S1). Identifying the anodic and cathodic current peaks occurring from the electrolysis of a redox-active solution, $[\text{Fe}(\text{CN})_6]^{3-/4-}$, the anodic and cathodic current values (I_a and I_c) were estimated from the respective peak intensities, and the charges (Q_a and Q_c) were calculated from the area encapsulated under the respective peaks. The results are given in Table S1 for all types of electrodes.




Figure 1. (A) Raman spectra of (a) nitrocellulose membrane (black), (b) carbon paste (red), (c) RGO-modified paper electrode (blue), (d) RGO-modified paper electrode after AuNP deposition (pink); (B) SEM images of (a) RGO and (b) AuNP/RGO-modified paper electrodes (scale: 5 μm). Gold deposition was performed with the chronoamperometric technique in aqueous solution of the HAuCl₄ gold precursor by applying −0.3 V for 10 min.




Figure 2. SEM image of RGO (scale: 5 μm).

The highest I_a and I_c were recorded by a AuNP/RGO-modified paper electrode (Table S1). Substantial increase in both I_a and I_c, compared with an RGO-modified paper electrode, confirmed that the role of AuNPs is to enhance the electrode conductivity by facilitating the electron transfer [43–45].

The electroactive surface area (A) of each electrode—unmodified paper electrode, RGO-modified paper electrode, and AuNP-decorated RGO-modified paper electrode—was calculated by using the Randles–Sevcik equation [46] (Equation (1)), where I_p is the peak current (I_a or I_c) in A, n is the number of transferred electrons, A is the surface area in

cm^2 , D is the diffusion coefficient in cm^2/s , C is the concentration of electroactive species in mol/cm^3 , and v is the scan rate in V/s .

$$I_p = 2.687 \times 10^5 \times n^{3/2} \times A \times D^{1/2} \times C \times v^{1/2} \quad (1)$$

The electroactive surface area of the paper electrodes was calculated based on Ia and found to be 0.020 cm^2 for the unmodified paper electrode, 0.026 cm^2 for the RGO-modified paper electrode, and 0.036 cm^2 for the AuNP-decorated RGO-modified paper electrode (shown in Table S1). An increase of about 80% was obtained at the electroactive surface area in the presence of a modification with AuNPs and RGO in comparison with the unmodified paper electrode due to the increase of the conductivity of the electrode based on the nature of the RGO nanomaterial and gold nanoparticles [47]. Furthermore, the AuNP/RGO-modified paper electrode exhibited about 39% increase in the electroactive surface area, confirming that the AuNP modification can enhance electroactivity, hence the sensitivity of the RGO-modified paper electrode.

3.2. Voltammetric Detection of miRNA-155 and miRNA-21 by a AuNP/RGO-Modified Paper Electrode

The detection of hybridization relies on the change of the oxidation signal of a redox $[\text{Fe}(\text{CN})_6]^{3-/4-}$ probe. The immobilization of a thiol-linked DNA probe onto the electrode leads to a decrease in peak current. This result suggests that the hindrance is caused by the negatively charged DNA probe, while preventing the diffusion of the redox probe $[\text{Fe}(\text{CN})_6]^{3-/4-}$ to the working electrode surface. The peak current was decreased after forming probe/miRNA target hybrids due to the presence of a more negatively charged DNA-miRNA hybrid at the electrode surface. The decrease at the peak current also indicates the forming perfect-match DNA probe/its complementary miRNA target hybrids [48].

All experiments for the detection of miRNA hybridization were efficiently carried out using a AuNP- and RGO-modified paper electrode to optimize the probe concentration, probe immobilization time, and hybridization time. The obtained results for the optimization studies are shown in Figures S2–S6 and Table S2.

Hybridization efficiency (HE%) is calculated as evidence of the probe and miRNA hybridization efficacy in order to determine optimum conditions [49].

$HE\% = \Delta I \times 100/I_{\text{probe}}$ represents the hybridization efficiency, where $\Delta I = I_{\text{hybrid}} - I_{\text{probe}}$.

All experiments related to the detection of miRNA-155 and miRNA-21 were further explored under optimum conditions of this study.

After the optimization studies, the analytical performance of the electrodes was tested through the detection of a miRNA-155 target at different concentrations in the range of $0.25\text{--}2.0 \text{ }\mu\text{g}/\text{mL}$. Accordingly, the voltammograms regarding the oxidation signals are shown in Figure 3A,B. The highest HE% is calculated and found to be 37.1% in the case of Probe-1 and $1.0 \text{ }\mu\text{g}/\text{mL}$ miRNA-155 target hybridization (see Table S3).

The identical procedure was applied for voltammetric detection of miRNA-21, which is another biomarker of non-small-cell lung carcinoma (NSCLC). Similarly, the analytical performance of the electrodes was tested through the detection of a miRNA-21 target at different concentrations in the range of $0.25\text{--}2.0 \text{ }\mu\text{g}/\text{mL}$. Accordingly, the voltammograms are shown in Figure 3C,D. The highest HE% is calculated and found to be 43.2% in the case of Probe-2 with $1.0 \text{ }\mu\text{g}/\text{mL}$ miRNA-21 target hybridization (see Table S4).

The detection limit (LOD) [50] was calculated to be $0.19 \text{ }\mu\text{g}/\text{mL}$ (25.71 nM , 128.0 fmol in $5.0 \text{ }\mu\text{L}$ sample) for miRNA-155 via linear fitting of the calibration curve with the equation $y = -10.63x + 27.05$ and $R^2 = 0.98$ (shown in Figure 4A). Similarly, the LOD of miRNA-21 was calculated to be $0.08 \text{ }\mu\text{g}/\text{mL}$ (12.0 nM , 60.0 fmol in $5.0 \text{ }\mu\text{L}$ sample) by fitting the calibration curve using the equation $y = -12.64x + 30$ and $R^2 = 0.99$ (Figure 4B). Additionally, the sensor sensitivity was estimated from the slope of the calibration curve, divided by the surface area of the AuNP/RGO-paper electrode, for miRNA-155 and miRNA-21, and found to be 295.3 and $351.1 \text{ }\mu\text{A}\cdot\text{mL}/\mu\text{g}\cdot\text{cm}^2$, respectively.




Figure 3. (A) Voltammograms representing the $[\text{Fe}(\text{CN})_6]^{3-/4-}$ oxidation signal obtained by (a) Probe-1 immobilized AuNP/RGO-paper electrode in the absence of miRNA-155 target, after hybridization of Probe-1 with miRNA-155 target at concentrations of (b) 0.25 $\mu\text{g/mL}$, (c) 0.5 $\mu\text{g/mL}$, (d) 0.75 $\mu\text{g/mL}$, (e) 1.0 $\mu\text{g/mL}$, (f) 1.5 $\mu\text{g/mL}$, and (g) 2.0 $\mu\text{g/mL}$. (B) The line graph based on the average $[\text{Fe}(\text{CN})_6]^{3-/4-}$ oxidation signal after hybridization between Probe-1 and miRNA-155 target with its various concentrations from 0 to 2.0 $\mu\text{g/mL}$ ($n = 3$). (C) Voltammograms representing the $[\text{Fe}(\text{CN})_6]^{3-/4-}$ oxidation signals obtained by (a) Probe-2 immobilized AuNP/RGO-paper electrode in the absence of miRNA-21 target, after hybridization of Probe-2 with miRNA-21 target at concentrations of (b) 0.25 $\mu\text{g/mL}$, (c) 0.5 $\mu\text{g/mL}$, (d) 0.75 $\mu\text{g/mL}$, (e) 1.0 $\mu\text{g/mL}$, (f) 1.5 $\mu\text{g/mL}$, and (g) 2.0 $\mu\text{g/mL}$. (D) The line graph based on the average $[\text{Fe}(\text{CN})_6]^{3-/4-}$ oxidation signal measured after hybridization between Probe-2 and miRNA-21 target with its various concentrations from 0 to 2.0 $\mu\text{g/mL}$ ($n = 3$).

3.3. Selectivity of the Assay on the Voltammetric Detection of miRNA-155 by the AuNP/RGO-Modified Paper Electrode

The selectivity of the assay was then investigated against other miRNAs; a single-base mismatch (MM) or noncomplementary (NC) ones and the results are given in Figure S7. In the absence of the target sequence, the average oxidation signal of $[\text{Fe}(\text{CN})_6]^{3-/4-}$ was measured to be $29.47 \pm 0.44 \mu\text{A}$. This signal decreased to $17.32 \pm 3.22 \mu\text{A}$ (RSD%, 18.64%, $n = 10$) after occurring the perfect-match Probe-1 and its target miRNA-155 hybrids (Figure S7). On the other hand, the average signal was obtained as $20.05 \pm 2.35 \mu\text{A}$ and $20.04 \pm 2.71 \mu\text{A}$ in the case of hybridization between Probe-1 and NC or MM, respectively (Figure S7). However, the oxidation peak current of $[\text{Fe}(\text{CN})_6]^{3-/4-}$ was measured to be $18.40 \pm 1.62 \mu\text{A}$ and $18.15 \pm 7.10 \mu\text{A}$ when hybridization was performed in the mixture samples consisting of target:NC (1:1) and target:MM (1:1), respectively (Figure S7). The highest decrease (i.e., 41.2%) at the oxidation signal of $[\text{Fe}(\text{CN})_6]^{3-/4-}$ was obtained in the case of a full-match hybridization in contrast to the ones obtained by NC or MM sequences (Table S5). Moreover, the standard deviations and RSD % values were high in the presence of NC or MM sequences due to the noneffective hybridization. Considering the number

of bases that are similar to the target sequence (4 base pairing with NC, 22 base pairing with MM, see supporting information), it is expected that the sensor developed exhibited a more selective behavior towards to the NC sequence than the MM sequence. In fact, the standard deviation and RSD% value obtained in the presence of NC were better than those obtained with MM. Hence, it can be concluded that the present assay offered a selective detection of miRNA even if the assay was examined in the mixture samples containing a miRNA target with other miRNA sequences, which differed one base from the target miRNA sequence or noncomplementary miRNA sequence (Table S5).




Figure 4. (A) Calibration plot for AuNP/RGO-paper electrode obtained after hybridization between Probe-1 and miRNA-155 target with its various concentrations from 0 to 1.0 $\mu\text{g/mL}$ ($n = 3$). (B) Calibration plot for the same electrode obtained after hybridization between Probe-2 and miRNA-21 target with its various concentrations from 0 to 1.0 $\mu\text{g/mL}$ ($n = 3$).

3.4. Selectivity of the Assay on the Detection of miRNA-21 by Differential Pulse Voltammetry Using an AuNP/RGO-Modified Paper Electrode

The selectivity of the assay was then investigated against NC or MM (Figure S8). The average oxidation signal of $[\text{Fe}(\text{CN})_6]^{3-/4-}$ was determined to be $17.00 \pm 3.17 \mu\text{A}$ (RSD%, 18.65%, $n = 2$) after forming the perfect-match Probe-2 and miRNA-21 target hybrids (Figure S8), whereas the average signal was measured to be $20.64 \pm 5.75 \mu\text{A}$ and $18.65 \pm 4.12 \mu\text{A}$ after the hybridization of Probe-2 with NC and MM, respectively (Figure S8). Hence, it can be concluded that the present assay offered a selective behavior even if the assay was formed from the mixture of the miRNA target and the oligonucleotides, which differed one base from target miRNA sequence or noncomplementary miRNA sequence (Table S6).

3.5. Characterization Studies of the Paper Electrode Modified with Gold Nanoparticle–Molybdenum Disulfide Nanosheets (AuNP/MoS₂)

The characterization of the AuNP- and MoS₂-modified paper electrode was achieved by Raman spectroscopy under the conditions indicated previously. Raman signals were obtained for the characterization of modified paper electrodes (Figure 5A). Three main Raman peaks in the wave number range of 300–500 cm^{−1} correspond to MoS₂ [51,52]. The peak at 381 cm^{−1} is attributed to the in-plane vibration of two S atoms and Mo (E¹_{2g}). The peak at 409 cm^{−1} is related to the out-plane vibration of S atoms (A¹_g). Another main MoS₂ peak at 452 cm^{−1} is due to the 2 LA mode. The obtained Raman spectra proved the existence of MoS₂ on the working electrode surface. As shown in Figure 5A, the SERS effect was observed after gold nanoparticle deposition on the modified surface. The signals of MoS₂ molecules were increased by gold deposition. This result is also evidence of gold deposition onto the surface.




Figure 5. (A) Raman spectra of paper electrode: (a) nitrocellulose membrane (black), (b) carbon paste (red), (c) MoS₂-modified paper electrode (blue), (d) MoS₂-modified paper electrode after AuNP deposition (red). (B) SEM images of (a) MoS₂-nanosheet-modified paper electrode and (b) MoS₂-modified paper electrode after gold electrodeposition (scale: 2 μm).

The MoS₂-nanosheet- and AuNP/MoS₂-modified paper electrodes were characterized by a Quanta 200 3D SEM. Figure 5B shows the deposition of bare MoS₂ nanosheets on the carbon-ink-modified NC paper electrode. After electrodeposition, the gold nanoparticles can clearly be seen on the MoS₂-modified paper electrode surface (Figure 5B). The diameter of AuNPs was measured to be 540 ± 140 nm. The SEM image in Figure 6 shows that the exfoliation process resulted in MoS₂ nanosheets with lateral dimensions of ~1 μm and a wide range of smaller nanosheets stacked on the larger ones.

The electrochemical characterization of the unmodified paper electrode, MoS₂-modified paper electrode, and AuNP deposition was performed by cyclic voltammetry (Figure S9). The charges (Q_a and Q_c) and currents (I_a (μA) and I_c (μA)) with the surface area of each electrode are shown in Table S7.

The electroactive surface area (A) was calculated according to I_a and found to be 0.020 cm² for the unmodified paper electrode, 0.021 cm² for the MoS₂-modified paper electrode, and 0.035 cm² for the AuNP/MoS₂-modified paper electrode (shown in Table S7). After AuNP/MoS₂ modification, the electroactive surface area of the AuNP/MoS₂-modified paper electrode was increased by about 75% compared with the unmodified one by means of a layered structure of MoS₂ nanosheets and the conductive nature of AuNPs.




Figure 6. SEM image of MoS₂ nanosheets (scale: 2 μm).

3.6. Voltammetric Detection of miRNA-155 and miRNA-21 by the AuNP- and MoS₂-Modified Paper Electrodes

All experiments were carried out by the AuNP- and MoS₂-modified paper electrodes for the optimization of the developed method, such as probe immobilization time and hybridization time. The obtained results are shown in Figures S10 and S11. Further experiments on miRNA-21 and miRNA-155 detection were carried out under optimum conditions in the present study.

The oxidation signals based on miRNA hybridization at different concentrations of miRNA-21 from 0.5 to 5.0 μg/mL were measured by the DPV technique. Figure 7A and Figure S12 show the representative voltammograms with the resulting line graph.

The LOD of miRNA-21 was calculated and found to be 0.36 μg/mL (51.68 nM, 258 fmol in 5.0 μL sample) using the equation $y = -2.77x + 23.39$ and $R^2 = 0.99$ (shown in Figure 7B) by AuNP/MoS₂-modified paper electrodes.

The calculated HE% values on hybridization with the miRNA-21 target are given in Table S8.

Similarly, the oxidation signals of miRNA-155 hybridization were measured voltammetrically at different concentrations of miRNA-155 from 1.0 to 4.0 μg/mL. Figure S13 shows the representative voltammograms with the line graph of the AuNP/MoS₂-modified paper electrodes. The highest HE% was calculated and found to be 32% in the presence of hybridization with a 2.0 μg/mL miRNA-155 target (see Table S9). The LOD [50] was also calculated and found to be 0.44 μg/mL (59.67 nM, 298 fmol in 5.0 μL sample) for miRNA-155 with the equation $y = -4.71x + 28.15$ and $R^2 = 0.97$ (shown in Figure S14).

Additionally, the sensitivity of the AuNP/MoS₂-modified paper electrode was estimated for miRNA-21 and miRNA-155 and found to be 79.1 and 134.6 μA·mL/μg·cm², respectively.

3.7. Selectivity of the Assay on the Detection of miRNA-155 by Differential Pulse Voltammetry Using the AuNP- and MoS₂-Modified Paper Electrodes

The selectivity of the assay was investigated against NC or MM (Figure S15). The average [Fe(CN)₆]^{3−/4−} oxidation signal was determined to be 19.74 ± 1.75 μA (RSD%, 8.88%, $n = 6$) after forming the perfect-match Probe-1/miRNA-155 target hybrids (Figure S15), whereas the average signal was recorded to be 30.77 ± 8.37 μA and 22.52 ± 2.80 μA after the hybridization of Probe-1 with NC and MM, respectively (Figure S15). Moreover, the developed paper-electrode-based DNA probe could identify its complementary target miRNAs with high selectivity in the samples containing NC or MM by measuring nearly the same signal in contrast to the perfect-match hybridization signal (Table S10). Hence, it

can be concluded that the developed assay offered a selective behavior even if the assay was formed from the mixture of the miRNA target and the oligonucleotides, which differed one base from the target miRNA sequence or noncomplementary miRNA sequence.




Figure 7. (A) Voltammograms representing the $[\text{Fe}(\text{CN})_6]^{3-/4-}$ oxidation signals obtained by (a) Probe-2 immobilized AuNP/MoS₂-paper electrode in the absence of miRNA-21 target, after hybridization of Probe-2 with miRNA-21 target at concentrations of (b) 0.5 $\mu\text{g/mL}$, (c) 1.0 $\mu\text{g/mL}$, (d) 1.5 $\mu\text{g/mL}$, (e) 2.0 $\mu\text{g/mL}$, (f) 2.5 $\mu\text{g/mL}$, (g) 3.0 $\mu\text{g/mL}$, and (h) 5.0 $\mu\text{g/mL}$. (B) The calibration plot for AuNP/MoS₂-paper electrode obtained after hybridization between Probe-2 and miRNA-21 target with its various concentrations from 0 to 3.0 $\mu\text{g/mL}$ ($n = 3$).

3.8. Selectivity of the Assay on the Detection of miRNA-21 by Differential Pulse Voltammetry Using the AuNP- and MoS₂-Modified Paper Electrodes

Similarly, the selectivity of the assay was investigated against NC or MM (Figure S16). The average $[\text{Fe}(\text{CN})_6]^{3-/4-}$ oxidation signal was recorded to be $16.20 \pm 3.12 \mu\text{A}$ (RSD%, 19.27%, $n = 8$) after occurring the perfect-match hybrid between Probe-2 and the miRNA-21 target (Figure S16), and there was a 29.73% decrease in comparison with the signal measured in the absence of the target. On the other hand, there were 17% and 6% increases and 6% and 5% decreases after the hybridization of Probe-2 with NC, MM, target:NC mixture, and target:MM mixture, respectively (Table S11). Since, the highest decrease at the $[\text{Fe}(\text{CN})_6]^{3-/4-}$ oxidation signal was obtained in the case of full-match hybridization in contrast to the ones obtained by NC or MM sequences, it can be concluded that the developed assay offered a selective behavior.

4. Conclusions

In this study, paper-based electrochemical biosensors were presented for sensitive detection of microRNA (i.e., miRNA-155 and miRNA-21) biomarkers related to early diagnosis of lung cancer for the first time. Hydrophobic barriers to creating electrode areas were constructed by wax printing, whereas the three-electrode system was fabricated by simple mask printing. The surface of the working electrode was modified using either

gold-nanoparticle-reduced graphene oxide or gold-nanoparticle–molybdenum disulfide nanosheets. The electroactive surface areas of AuNP/RGO and AuNP/MoS₂-modified paper electrodes (about 80% and 75%, respectively) were increased with respect to unmodified ones. The resulting paper-based biosensors exhibited good reproducibility by the incorporation of unique properties of RGO and MoS₂ nanosheets. Additionally, AuNPs played an excellent role in the signal amplification.

Here, the voltammetric analysis of miRNA-155 and miRNA-21 resulted in a relatively shorter detection time in comparison with earlier studies related to biosensors (Table 1). The entire assay performed at room temperature, including electrode modification and miRNA detection, was completed in 35 min. A single droplet (5.0 µL) of a sample was enough to cover the entire working electrode area, which enabled analysis in low sample volumes. Barring a few exceptions, the sample volumes used in previous works are in the range of 5–100 µL. Therefore, the sample volume of our assay is one of the lowest volumes among the studies summarized in Table 1. The LODs of miRNA-21 were calculated to be 12.00 and 51.68 nM using a AuNP/RGO-modified paper electrode and a AuNP/MoS₂-modified paper electrode, respectively. On the other hand, the LODs of miRNA-155 were found to be 25.71 and 59.67 nM using a AuNP/RGO-modified paper electrode and a AuNP/MoS₂-modified paper electrode, respectively. In contrast to the results obtained by the AuNP/MoS₂-modified paper electrode, the AuNP/RGO-modified paper electrode performed miRNA detection with more sensitive results. Overall, the studies indicate that our proposed assay with nanosheet-modified paper electrodes detected miRNA hybridization accurately in contrast to one-base mismatch miRNA or noncomplementary miRNA. The proposed assay offers some advantages over earlier reports on miRNA detection (summarized in Table 1) in terms of ease of use, short assay time (35 min), and low cost per analysis. Additionally, it is important to note that our method simplifies the miRNA detection assay by avoiding the complex chemistries (i.e., cleaning of the electrode surface, formation of a self-assembled monolayer, usage of a nanoparticle-attached DNA probe) in sensor fabrication steps in comparison to earlier reports [53,54].

Table 1. Comparison of different electrochemical biosensors for the detection of miRNA-155 and miRNA-21.

miRNA	Electrode	Method	Analysis Time	Sample Volume	Concentration Range	DL	Reference
miRNA-107	SPGE	DPV	75 min	30 µL	5 fM–5 pM	10 fM	[55]
	Au-NPFe ₂ O ₃ NC/SPCE	CC	45 min	-	100 aM–1 nM	100 aM	[56]
	MoS ₂ /Thi/AuNPs nanocomposite/GCE	SWV	18 h	5 µL	1 pM–10 nM	0.26 pM	[57]
	Au@NPFe ₂ O ₃ NC/GCE	CC	-	-	100 fM–1 µM	100 fM	[54]
	AuNPs/ITO	ASV	17 h	100 µL	2.5 fM–25 nM	0.12 fM	[58]
miRNA-21	AuNPs/GCE	DPV	3.5 h	40 µL	100 aM–1 nM	78.0 aM	[59]
	AuNPs@MoS ₂ /GCE	DPV EIS	17 h	20 µL	10 fM–1 nM	0.78 fM 0.45 fM	[60]
	MWCNTs@GONRs/AuNPs/GCE	DPV	14 h	6 µL	0.1 fM–0.1 nM	0.034 fM	[61]
	AuE	DPV	14 h	2 µL	0.1 fM–1 nM	0.04 fM	[62]
	AuNPs@MoS ₂ /SPGE AuNPs@/SPGE	CC	22 h	10 µL	100 aM–1 pM 10 fM–10 pM	100.0 aM 10.0 fM	[49]
	AuNPs/RGO/PE AuNPs/MoS ₂ /PE	DPV	35 min	5 µL	37.5 nM–150 nM 71.7 nM–430.2 nM	12 nM 51.7 nM	This study
	AuE	SWV	8 h	-	0.5 pM–0.1 µM	0.13 fM	[63]
	GO/Au/GCE	CV, DPV	21 h	10 µL	0.8 fM–1 nM	0.37 fM	[64]
miRNA-155	nano-Pd/Thi/GCE	CV	17 h	20 µL	5.6 pM–5.6 µM	1.87 pM	[65]
	AuNRs/GO/GCE	DPV	4 h	5 µL	2 fM–8 pM	0.6 fM	[66]
	AuNPs/RGO/PE AuNPs/MoS ₂ /PE	DPV	35 min	5 µL	33.8 nM–135.3 nM 135.6 nM–406.8 nM	25.7 nM 59.7 nM	This study

Abbreviations: MWCNT: multiwalled carbon nanotube, GCE: glassy carbon electrode, AuNRs: gold nanorods, AuE: gold electrode, ITO: indium tin oxide, GONRs: graphene oxide nanoribbons, SPGE: screen-printed gold electrode, DPV: differential pulse voltammetry, CC: chronocoulometry, SWV: square wave voltammetry, ASV: stripping voltammetry, EIS: electrochemical impedance spectroscopy.

Supplementary Materials: The following are available online at <https://www.mdpi.com/article/10.3390/bios11070236/s1>, Scheme S1. The schematic illustration of RGO/MoS₂-modified paper electrode assembly fabrication and Probe/miRNA assembling; Figure S1. CVs recorded in optimum conditions by using (a) unmodified paper electrode, (b) RGO-modified paper electrode, (c) after activation of RGO-modified paper electrode using covalent agents, (d) after electrodeposition of AuNPs onto the surface of chemically activated and RGO-modified paper electrode in the presence of 50.0 mM potassium ferricyanide in 100.0 mM KCl; Figure S2. The images of (a) 2.5 mM, (b) 5.0 mM, (c) 10.0 mM, (d) 15.0 mM HAuCl₄ deposited RGO-modified paper electrode; Figure S3. The images of AuNPs/RGO-modified paper electrode after deposition of 5.0 mM HAuCl₄ during (a) 5 min, (b) 7 min, (c) 10 min, (d) 15 min; Figure S4. (A) DPVs (A') histograms representing (a) AuNPs/RGO-modified paper electrode, (b) 0.5 µg/mL Probe-1 immobilized AuNPs/RGO-modified paper electrode, after hybridization of 0.5 µg/mL miRNA 155 Probe-1 with (c) 2.0 µg/mL miRNA-155 target (n = 2). (B) DPVs (B') histograms representing (d) 1.0 µg/mL miRNA 155 Probe-1 immobilized AuNPs/RGO-modified paper electrode, after hybridization of 1.0 µg/mL Probe-1 with (e) 2.0 µg/mL miRNA-155 target (n = 2). (C) DPVs (C') histograms representing (f) 2.0 µg/mL Probe-1 immobilized AuNPs/RGO-modified paper electrode, after hybridization of 2.0 µg/mL Probe-1 with (g) 2.0 µg/mL miRNA-155 target (n = 2); Figure S5. (A) DPVs (B) histograms representing 1.0 mM [Fe(CN)₆]^{3−/4−} oxidation signal obtained by (a) AuNPs/RGO-modified paper electrode, after immobilization of 0.5 µg/mL Probe-1 during (b) 10 min (c) 30 min onto the surface of AuNPs/RGO-modified paper electrode (n = 3); Figure S6. (A) DPVs, (B) histograms representing the [Fe(CN)₆]^{3−/4−} oxidation signal obtained by (a) AuNPs/RGO-modified paper electrode, (b) 0.5 µg/mL Probe-1 immobilized AuNPs/RGO-modified paper electrode, after the hybridization of Probe-1 with miRNA-155 target during (c) 5 min, (d) 15 min (n = 4); Figure S7. (A) Voltammograms representing the [Fe(CN)₆]^{3−/4−} oxidation signal obtained by (a) Probe-1 immobilized AuNPs/RGO-modified paper electrode in the absence of miRNA-155 target, after hybridization of Probe-1 with (b) miRNA-155 target, (c) NC, and (d) MM, individually. (B) Voltammograms representing the [Fe(CN)₆]^{3−/4−} oxidation signal obtained by (a) Probe-1 immobilized AuNPs/RGO-modified paper electrode in the absence of miRNA-155 target, after hybridization of Probe-1 (b) with only miRNA-155 target, (c) in target:NC (1:1) mixture, and (d) in target:MM (1:1) mixture; Figure S8. (A) Voltammograms representing the [Fe(CN)₆]^{3−/4−} oxidation signal obtained by (a) Probe-2 immobilized AuNPs/RGO-modified paper electrode in the absence of miRNA-21 target, after hybridization of Probe-2 with (b) miRNA-21 target, (c) NC, and (d) MM, individually. (B) Voltammograms representing the [Fe(CN)₆]^{3−/4−} oxidation signal obtained by (a') Probe-2 immobilized AuNPs/RGO-modified paper electrode in the absence of miRNA-21 target, after hybridization of Probe-2 (b') with only miRNA-21 target, (c') in target:NC (1:1) mixture, and (d') in target:MM (1:1) mixture; Figure S9. CVs recorded in optimum conditions by (a) unmodified paper electrode, (b) MoS₂-modified paper electrode, (c) AuNPs electrodeposited MoS₂-modified paper electrodes in the presence of 50.0 mM potassium ferricyanide in 100.0 mM KCl; Figure S10. (A) DPVs (B) histograms representing (a) AuNPs/MoS₂-modified paper electrode, (b) 0.5 µg/mL DNA probe immobilized AuNPs/MoS₂-modified paper electrode, (c) after hybridization of 0.5 µg/mL DNA probe with 1.0 µg/mL miRNA-155 target (n = 2). (C) DPVs (D) histograms representing (a) AuNPs/MoS₂-modified paper electrode, (b) 1.0 µg/mL DNA probe immobilized AuNPs/MoS₂-modified paper electrode, (c) after hybridization of 1.0 µg/mL DNA probe with 1.0 µg/mL miRNA-155 target (n = 2); Figure S11. (A) DPVs, (B) histograms representing the [Fe(CN)₆]^{3−/4−} oxidation signal obtained by (a) AuNPs/MoS₂-modified paper electrode, (b) 0.5 µg/mL Probe-1 immobilized AuNPs/MoS₂-modified paper electrode, after the hybridization of Probe-1 with miRNA-155 target during (c) 5 min, (d) 15 min (n = 3); Figure S12. The line graph based on the average [Fe(CN)₆]^{3−/4−} oxidation signal measured after hybridization between Probe-2 and miRNA-21 target with its various concentrations from 0 to 5.0 µg/mL (n = 3); Figure S13. (A) Voltammograms representing the [Fe(CN)₆]^{3−/4−} oxidation signals obtained by (a) Probe-1 immobilized AuNPs/MoS₂-paper electrode, after hybridization of Probe-1 with miRNA-155 target at the concentrations of (b) 1.0 µg/mL, (c) 2.0 µg/mL, (d) 3.0 µg/mL, (e) 4.0 µg/mL. (B) The line graph based on the average [Fe(CN)₆]^{3−/4−} oxidation signal after hybridization between Probe-1 and miRNA-155 target with its various concentrations from 0 to 4.0 µg/mL (n = 3); Figure S14. The calibration plot obtained after hybridization between Probe-1 and miRNA-155 target with its various concentrations from 0 to 2.0 µg/mL (n = 3); Figure S15. (A) Voltammograms representing the [Fe(CN)₆]^{3−/4−} oxidation signal obtained by (a) Probe-1 immobilized AuNPs/MoS₂-modified paper electrode in the absence of miRNA-155 target, after hybridization of Probe-1 with (b) miRNA-155 tar-

get, (c) NC, and (d) MM, individually. (B) Voltammograms representing the $[\text{Fe}(\text{CN})_6]^{3-/4-}$ oxidation signal obtained by (a) Probe-1 immobilized AuNPs/ MoS_2 -modified paper electrode in the absence of miRNA-155 target, after hybridization of Probe-1 (b) with only miRNA-155 target, (c) in target:NC (1:1) mixture, and (d) in target:MM (1:1) mixture; Figure S16. (A) Voltammograms representing the $[\text{Fe}(\text{CN})_6]^{3-/4-}$ oxidation signal obtained by (a) Probe-2 immobilized AuNPs/ MoS_2 -modified paper electrode in the absence of miRNA-21 target, after hybridization of Probe-2 with, (b) miRNA-21 target, (c) NC, and (d) MM, individually. (B) Voltammograms representing the $[\text{Fe}(\text{CN})_6]^{3-/4-}$ oxidation signal obtained by (a) Probe-2 immobilized AuNPs/ MoS_2 -modified paper electrode in the absence of miRNA-21 target, after hybridization of Probe-2 (b) with only miRNA-21 target, (c) target:NC (1:1) mixture, and (d) in target:MM (1:1) mixture; Table S1. The anodic current I_a (μA) and the cathodic current I_c (μA), the relative charge, Q_a and Q_c of $[\text{Fe}(\text{CN})_6]^{3-/4-}$ measured by unmodified, RGO modified, after activation of RGO-modified paper electrode using covalent agents and AuNPs/RGO-modified paper electrode.; Table S2. The oxidation signal of $[\text{Fe}(\text{CN})_6]^{3-/4-}$ measured before/after 0.5 $\mu\text{g/mL}$ DNA probe immobilization onto the surface of AuNPs/RGO-modified paper electrode during 10 and 30 min and HE% values; Table S3. The oxidation signal of $[\text{Fe}(\text{CN})_6]^{3-/4-}$ measured before/after hybridization of Probe-1 and miRNA-155 target in its different concentrations ($n = 3$) and HE% values; Table S4. The oxidation signal of $[\text{Fe}(\text{CN})_6]^{3-/4-}$ and decrease % at the signal after hybridization of Probe-2 and miRNA-21 target in its different concentrations ($n = 3$); Table S5. The average $[\text{Fe}(\text{CN})_6]^{3-/4-}$ oxidation signals measured before and after hybridization of Probe-1 with miRNA-155 target, NC, MM, the mixture sample containing target:NC (1:1) or the mixture sample containing target:MM (1:1). HE% calculated according to the oxidation signals obtained after hybridization; Table S6. The average $[\text{Fe}(\text{CN})_6]^{3-/4-}$ oxidation signals measured before and after hybridization of Probe-2 with miRNA-21 target, NC, MM, the mixture sample containing target:NC (1:1) or the mixture sample containing target:MM (1:1). HE % calculated according to the oxidation signals obtained after hybridization; Table S7. The anodic current I_a (μA) and the cathodic current I_c (μA), the relative charge, Q_a and Q_c of $[\text{Fe}(\text{CN})_6]^{3-/4-}$ measured by unmodified, MoS_2 modified and AuNPs/ MoS_2 -modified paper electrode; Table S8. The oxidation signal of $[\text{Fe}(\text{CN})_6]^{3-/4-}$ and HE% values that calculated after hybridization of Probe-2 and miRNA-21 target in its different concentrations ($n = 3$); Table S9. The oxidation signal of $[\text{Fe}(\text{CN})_6]^{3-/4-}$ and HE% values that calculated after hybridization of Probe-1 and miRNA-155 target in its different concentrations ($n = 3$); Table S10. The average $[\text{Fe}(\text{CN})_6]^{3-/4-}$ oxidation signals ($n = 2$) measured before and after hybridization of Probe-1 with miRNA-155 target, NC, MM, the mixture sample containing target:NC (1:1) or the mixture sample containing target:MM (1:1). HE% calculated according to the oxidation signals obtained after hybridization; Table S11. The average $[\text{Fe}(\text{CN})_6]^{3-/4-}$ oxidation signals ($n = 2$) measured before and after hybridization of Probe-2 with miRNA-21 target, NC, MM, the mixture sample containing target:NC (1:1) or the mixture sample containing target:MM (1:1). HE% calculated according to the oxidation signals obtained after hybridization.

Author Contributions: Conceptualization, A.E., U.T. and P.P.; methodology, A.E.; U.T. and P.P.; investigation, H.T., E.Y., E.E., A.G., J.B.; data curation, H.T., E.Y., E.E., A.G., J.B.; writing—original draft preparation, H.T., E.E., A.E., U.T. and P.P.; writing—review and editing, A.E., U.T. and P.P.; supervision, A.E., U.T. and P.P.; project administration, A.E. and P.P.; funding acquisition, A.E. and P.P. All authors have read and agreed to the published version of the manuscript.

Funding: This project was supported by the Newton-Katip Celebi funding program, and received a financial support from the Turkish Scientific and Technological Research Council (TUBITAK; Project no. 215Z702), the British Council (Newton Fund, Institutional Links, Ref: 216182787) and Invest Northern Ireland under a Biodevices grant, Ref. RD0714186.

Acknowledgments: This project was supported by the Newton-Katip Celebi funding program, and authors acknowledge financial support from the Turkish Scientific and Technological Research Council (TUBITAK; Project no. 215Z702) and the British Council (Newton Fund, Institutional Links, Ref: 216182787). A.E. would also like to express her gratitude to the Turkish Academy of Sciences (TUBA) as a principal member for its partial support. E.E. and E.Y., PhD and master students, respectively, acknowledge a project scholarship (TUBITAK Project no. 215Z702). Authors also acknowledge to helpful discussion of Yildiz Uludag as the project consultant during project (TUBITAK; Project no. 215Z702). P.P. acknowledges support from Invest Northern Ireland under a Biodevices grant, Ref. RD0714186.

Conflicts of Interest: The authors declare no conflict of interest.

References

- Yagoda, H. Applications of confined spot tests in analytical chemistry. *Ind. Eng. Chem. Anal. Ed.* **1937**, *9*, 79–82. [\[CrossRef\]](#)
- Muller, R.H.; Clegg, D.L. Automatic paper chromatography. *Anal. Chem.* **1949**, *21*, 1123–1125. [\[CrossRef\]](#)
- Comer, J.P. Semiquantitative specific test paper for glucose in urine. *Anal. Chem.* **1956**, *28*, 1748–1750. [\[CrossRef\]](#)
- Martinez, A.W.; Phillips, S.T.; Butte, M.J.; Whitesides, G.M. Patterned paper as a platform for inexpensive, low-volume, portable bioassays. *Angew. Chem. Int. Ed.* **2007**, *46*, 1318–1320. [\[CrossRef\]](#)
- Fu, L.M.; Wang, Y.N. Detection methods and applications of microfluidic paper-based analytical devices. *TrAC Trends Anal. Chem.* **2018**, *107*, 196–211. [\[CrossRef\]](#)
- Xia, Y.; Si, J.; Li, Z. Fabrication techniques for microfluidic paper-based analytical devices and their applications for biological testing: A review. *Biosens. Bioelectron.* **2016**, *77*, 774–789. [\[CrossRef\]](#) [\[PubMed\]](#)
- Nilghaz, A.; Guan, L.; Tan, W.; Shen, W. Advances of Paper-Based Microfluidics for Diagnostics—The Original Motivation and Current Status. *ACS Sens.* **2016**, *1*, 1382–1393. [\[CrossRef\]](#)
- Torul, H.; Ciftci, H.; Cetin, D.; Suludere, Z.; Boyaci, I.H.; Tamer, U. Paper membrane-based SERS platform for the determination of glucose in blood samples. *Anal. Bioanal. Chem.* **2015**, *407*, 8243–8251. [\[CrossRef\]](#) [\[PubMed\]](#)
- Dungchai, W.; Chailapakul, O.; Henry, C.S. Electrochemical detection for paper-based microfluidics. *Anal. Chem.* **2009**, *81*, 5821–5826. [\[CrossRef\]](#) [\[PubMed\]](#)
- Zhu, X.; Ino, K.; Lin, Z.; Shiku, H.; Chen, G.; Matsue, T. Amperometric detection of DNA hybridization using a multi-point, addressable electrochemical device. *Sens. Actuators B Chem.* **2011**, *160*, 923–928. [\[CrossRef\]](#)
- Nie, Z.; Nijhuis, C.A.; Gong, J.; Chen, X.; Kumachev, A.; Martinez, A.W.; Narovlyansky, M.; Whitesides, G.M. Electrochemical sensing in paper-based microfluidic devices. *Lab Chip* **2010**, *10*, 477–483. [\[CrossRef\]](#) [\[PubMed\]](#)
- Sjöberg, P.; Määttä, A.; Vanamo, U.; Novell, M.; Ihalainen, P.; Andrade, F.J.; Bobacka, J.; Peltonen, J. Paper-based potentiometric ion sensors constructed on ink-jet printed gold electrodes. *Sens. Actuators B Chem.* **2016**, *224*, 325–332. [\[CrossRef\]](#)
- Sun, L.J.; Xie, Y.; Yan, Y.F.; Yang, H.; Gu, H.Y.; Bao, N. Paper-based analytical devices for direct electrochemical detection of free IAA and SA in plant samples with the weight of several milligrams. *Sens. Actuators B Chem.* **2017**, *247*, 336–342. [\[CrossRef\]](#)
- Wu, J.; Chen, Q.; Liu, W.; He, Z.; Lin, J.M. Recent advances in microfluidic 3D cellular scaffolds for drug assays. *TrAC Trends Anal. Chem.* **2017**, *87*, 19–31. [\[CrossRef\]](#)
- Guzman, J.M.C.C.; Tayo, L.L.; Liu, C.C.; Wang, Y.N.; Fu, L.M. Rapid microfluidic paper-based platform for low concentration formaldehyde detection. *Sens. Actuators B Chem.* **2018**, *255*, 3623–3629. [\[CrossRef\]](#)
- Tian, T.; Liu, H.; Li, L.; Yu, J.; Ge, S.; Song, X.; Yan, M. Paper-based biosensor for noninvasive detection of epidermal growth factor receptor mutations in non-small cell lung cancer patients. *Sens. Actuators B Chem.* **2017**, *251*, 440–445. [\[CrossRef\]](#)
- Sriram, G.; Bhat, M.P.; Patil, P.; Uthappa, U.T.; Jung, H.Y.; Altalhi, T.; Kumeria, T.; Aminabhavi, T.M.; Pai, R.K.; Madhuprasad, et al. Paper-based microfluidic analytical devices for colorimetric detection of toxic ions: A review. *TrAC Trends Anal. Chem.* **2017**, *93*, 212–227. [\[CrossRef\]](#)
- Arduini, F.; Micheli, L.; Moscone, D.; Palleschi, G.; Piermarini, S.; Ricci, F.; Volpe, G. Electrochemical biosensors based on nanomodified screen-printed electrodes: Recent applications in clinical analysis. *TrAC Trends Anal. Chem.* **2016**, *79*, 114–126. [\[CrossRef\]](#)
- Mak, W.C.; Beni, V.; Turner, A.P.F. Lateral-flow technology: From visual to instrumental. *TrAC Trends Anal. Chem.* **2016**, *79*, 297–305. [\[CrossRef\]](#)
- Zarei, M. Portable biosensing devices for point-of-care diagnostics: Recent developments and applications. *TrAC Trends Anal. Chem.* **2017**, *91*, 26–41. [\[CrossRef\]](#)
- Pol, R.; Céspedes, F.; Gabriel, D.; Baeza, M. Microfluidic lab-on-a-chip platforms for environmental monitoring. *TrAC Trends Anal. Chem.* **2017**, *95*, 62–68. [\[CrossRef\]](#)
- Lee, Y.S.; Dutta, A. MicroRNAs in Cancer. *Annu. Rev. Pathol. Mech. Dis.* **2009**, *4*, 199–227. [\[CrossRef\]](#) [\[PubMed\]](#)
- Lee, R.; Feinbaum, R.; Ambros, V. The C. elegans Heterochronic Gene lin-4 Encodes Small RNAs with Antisense Complementarity to lin-14. *Cell* **1993**, *75*, 843–854. [\[CrossRef\]](#)
- MacFarlane, L.-A.; Murphy, P.R. MicroRNA: Biogenesis, Function and Role in Cancer. *Curr. Genom.* **2010**, *11*, 537–561. [\[CrossRef\]](#)
- Islam, M.N.; Masud, M.K.; Haque, M.H.; Al Hossain, M.S.; Yamauchi, Y.; Nguyen, N.-T.; Shiddiky, M.J.A. RNA Biomarkers: Diagnostic and Prognostic Potentials and Recent Developments of Electrochemical Biosensors. *Small Methods* **2017**, *1*, 1700131. [\[CrossRef\]](#)
- Negrini, M.; Ferracin, M.; Sabbioni, S.; Croce, C.M. MicroRNAs in human cancer: From research to therapy. *J. Cell Sci.* **2007**, *120*, 1833–1840. [\[CrossRef\]](#)
- Eis, P.S.; Tam, W.; Sun, L.; Chadburn, A.; Li, Z.; Gomez, M.F.; Lund, E.; Dahlberg, J.E. Accumulation of miR-155 and BIC RNA in human B cell lymphomas. *Proc. Natl. Acad. Sci. USA* **2005**, *102*, 3627–3632. [\[CrossRef\]](#)
- Chan, J.A.; Krichevsky, A.M.; Kosik, K.S. MicroRNA-21 is an antiapoptotic factor in human glioblastoma cells. *Cancer Res.* **2005**, *65*, 6029–6033. [\[CrossRef\]](#)
- Kilic, T.; Erdem, A.; Ozsoz, M.; Carrara, S. microRNA biosensors: Opportunities and challenges among conventional and commercially available techniques. *Biosens. Bioelectron.* **2018**, *99*, 525–546. [\[CrossRef\]](#)

30. Erdem, A.; Congur, G.; Eksin, E. Multi channel screen printed array of electrodes for enzyme-linked voltammetric detection of MicroRNAs. *Sens. Actuators B Chem.* **2013**, *188*, 1089–1095. [\[CrossRef\]](#)
31. Isin, D.; Eksin, E.; Erdem, A. Graphene oxide modified single-use electrodes and their application for voltammetric miRNA analysis. *Mater. Sci. Eng. C* **2017**, *75*, 1242–1249. [\[CrossRef\]](#)
32. Bartosik, M.; Trefulka, M.; Hrstka, R.; Vojtesek, B.; Palecek, E. Os(VI)bipy-based electrochemical assay for detection of specific microRNAs as potential cancer biomarkers. *Electrochem. Commun.* **2013**, *33*, 55–58. [\[CrossRef\]](#)
33. Masud, M.K.; Umer, M.; Hossain, M.S.A.; Yamauchi, Y.; Nguyen, N.T.; Shiddiky, M.J.A. Nanoarchitecture Frameworks for Electrochemical miRNA Detection. *Trends Biochem. Sci.* **2019**, *44*, 433–452. [\[CrossRef\]](#)
34. Sage, A.T.; Besant, J.D.; Lam, B.; Sargent, E.H.; Kelley, S.O. Ultrasensitive electrochemical biomolecular detection using nanostructured microelectrodes. *Acc. Chem. Res.* **2014**, *47*, 2417–2425. [\[CrossRef\]](#)
35. Lee, S.H.; Lee, J.H.; Tran, V.K.; Ko, E.; Park, C.H.; Chung, W.S.; Seong, G.H. Determination of acetaminophen using functional paper-based electrochemical devices. *Sens. Actuators B Chem.* **2016**, *232*, 514–522. [\[CrossRef\]](#)
36. Ding, R.; Fiedoruk-Pogrebniak, M.; Pokrzywnicka, M.; Koncki, R.; Bobacka, J.; Lisak, G. Solid reference electrode integrated with paper-based microfluidics for potentiometric ion sensing. *Sens. Actuators B Chem.* **2020**, *323*, 128680. [\[CrossRef\]](#)
37. Yukird, J.; Soum, V.; Kwon, O.S.; Shin, K.; Chailapakul, O.; Rodthongkum, N. 3D paper-based microfluidic device: A novel dual-detection platform of bisphenol A. *Analyst* **2020**, *145*, 1491–1498. [\[CrossRef\]](#) [\[PubMed\]](#)
38. Eksin, E.; Torul, H.; Yarali, E.; Tamer, U.; Papakonstantinou, P.; Erdem, A. Paper-based electrode assemble for impedimetric detection of miRNA. *Talanta* **2021**, *225*, 122043. [\[CrossRef\]](#) [\[PubMed\]](#)
39. Benson, J.; Li, M.; Wang, S.; Wang, P.; Papakonstantinou, P. Electrocatalytic Hydrogen Evolution Reaction on Edges of a Few Layer Molybdenum Disulfide Nanodots. *ACS Appl. Mater. Interfaces* **2015**, *7*, 14113–14122. [\[CrossRef\]](#)
40. Hou, Z.Q.; Wang, Z.Y.; Yang, L.X.; Yang, Z.G. Nitrogen-doped reduced graphene oxide intertwined with V₂O₃ nanoflakes as self-supported electrodes for flexible all-solid-state supercapacitors. *RSC Adv.* **2017**, *7*, 25732–25739. [\[CrossRef\]](#)
41. Pal, A.; Kar, S.; Debnath, A.K.; Aswal, D.K.; Bindal, R.C.; Tewari, P.K. Reinforcement of nanostructured reduced graphene oxide: A facile approach to develop high-performance nanocomposite ultrafiltration membranes minimizing the trade-off between flux and selectivity. *RSC Adv.* **2015**, *5*, 46801–46816. [\[CrossRef\]](#)
42. Jasuja, K.; Berry, V. Implantation and growth of dendritic gold nanostructures on graphene derivatives: Electrical property tailoring and Raman enhancement. *ACS Nano* **2009**, *3*, 2358–2366. [\[CrossRef\]](#)
43. Sun, W.; Wang, D.; Zhang, Y.Y.; Ju, X.M.; Yang, H.X.; Chen, Y.X.; Sun, Z.F. Electrodeposited graphene and gold nanoparticle modified carbon ionic liquid electrode for sensitive detection of rutin. *Chin. J. Anal. Chem.* **2013**, *41*, 709–713. [\[CrossRef\]](#)
44. Wu, B.; Hou, L.; Zhang, T.; Han, Y.; Kong, C. A molecularly imprinted electrochemical sensor based on a gold nanoparticle/carbon nanotube hybrid material for the sensitive detection of isoniazid. *Anal. Methods* **2015**, *7*, 9121–9129. [\[CrossRef\]](#)
45. Zhang, F.; Yuan, Y.; Zheng, Y.; Wang, H.; Liu, T.; Hou, S. A glassy carbon electrode modified with gold nanoparticle-encapsulated graphene oxide hollow microspheres for voltammetric sensing of nitrite. *Microchim. Acta* **2017**, *184*, 1565–1572. [\[CrossRef\]](#)
46. Cummings, T.E.; Elving, P.J. Determination of the Electrochemically Effective Electrode Area. *Anal. Chem.* **1978**, *50*, 480–488. [\[CrossRef\]](#)
47. Huang, K.J.; Liu, Y.J.; Wang, H.B.; Wang, Y.Y.; Liu, Y.M. Sub-femtomolar DNA detection based on layered molybdenum disulfide/multi-walled carbon nanotube composites, au nanoparticle and enzyme multiple signal amplification. *Biosens. Bioelectron.* **2014**, *55*, 195–202. [\[CrossRef\]](#)
48. Zhang, W.; Yang, T.; Li, X.; Wang, D.; Jiao, K. Conductive architecture of Fe₂O₃ microspheres/self-doped polyaniline nanofibers on carbon ionic liquid electrode for impedance sensing of DNA hybridization. *Biosens. Bioelectron.* **2009**, *25*, 428–434. [\[CrossRef\]](#)
49. Ganguly, A.; Benson, J.; Papakonstantinou, P. Sensitive chronocoulometric detection of miRNA at screen-printed electrodes modified by gold-decorated MoS₂ nanosheets. *ACS Appl. Bio Mater.* **2018**, *1*, 1184–1194. [\[CrossRef\]](#)
50. Miller, J.N.; Miller, J.C. *Statistics and Chemometrics for Analytical Chemistry*, 6th ed.; Pearson: Harlow, UK, 2010; ISBN 0273730428/9780273730422.
51. Sahu, T.S.; Mitra, S. Exfoliated MoS₂ Sheets and Reduced Graphene Oxide-An Excellent and Fast Anode for Sodium-ion Battery. *Sci. Rep.* **2015**, *5*, 1–13. [\[CrossRef\]](#)
52. Chen, L.; Chen, F.; Tronganh, N.; Lu, M.; Jiang, Y.; Gao, Y.; Jiao, Z.; Cheng, L.; Zhao, B. MoS₂/graphene nanocomposite with enlarged interlayer distance as a high performance anode material for lithium-ion battery. *J. Mater. Res.* **2016**, *31*, 3151–3160. [\[CrossRef\]](#)
53. Tavallaie, R.; McCarroll, J.; Le Grand, M.; Ariotti, N.; Schuhmann, W.; Bakker, E.; Tilley, R.D.; Hibbert, D.B.; Kavallaris, M.; Gooding, J.J. Nucleic acid hybridization on an electrically reconfigurable network of gold-coated magnetic nanoparticles enables microRNA detection in blood. *Nat. Nanotechnol.* **2018**, *13*, 1066–1071. [\[CrossRef\]](#) [\[PubMed\]](#)
54. Kamal Masud, M.; Islam, M.N.; Haque, M.H.; Tanaka, S.; Gopalan, V.; Alici, G.; Nguyen, N.T.; Lam, A.K.; Hossain, M.S.A.; Yamauchi, Y.; et al. Gold-loaded nanoporous superparamagnetic nanocubes for catalytic signal amplification in detecting miRNA. *Chem. Commun.* **2017**, *53*, 8231–8234. [\[CrossRef\]](#) [\[PubMed\]](#)
55. Koo, K.M.; Carrascosa, L.G.; Shiddiky, M.J.A.; Trau, M. Poly(A) Extensions of miRNAs for Amplification-Free Electrochemical Detection on Screen-Printed Gold Electrodes. *Anal. Chem.* **2016**, *88*, 2000–2005. [\[CrossRef\]](#)

56. Islam, M.N.; Masud, M.K.; Nguyen, N.T.; Gopalan, V.; Alamri, H.R.; Alothman, Z.A.; Al Hossain, M.S.; Yamauchi, Y.; Lam, A.K.Y.; Shiddiky, M.J.A. Gold-loaded nanoporous ferric oxide nanocubes for electrocatalytic detection of microRNA at attomolar level. *Biosens. Bioelectron.* **2018**, *101*, 275–281. [[CrossRef](#)]
57. Zhu, D.; Liu, W.; Zhao, D.; Hao, Q.; Li, J.; Huang, J.; Shi, J.; Chao, J.; Su, S.; Wang, L. Label-Free Electrochemical Sensing Platform for MicroRNA-21 Detection Using Thionine and Gold Nanoparticles Co-Functionalized MoS₂ Nanosheet. *ACS Appl. Mater. Interfaces* **2017**, *9*, 35597–35603. [[CrossRef](#)]
58. Guo, W.J.; Wu, Z.; Yang, X.Y.; Pang, D.W.; Zhang, Z.L. Ultrasensitive electrochemical detection of microRNA-21 with wide linear dynamic range based on dual signal amplification. *Biosens. Bioelectron.* **2019**, *131*, 267–273. [[CrossRef](#)] [[PubMed](#)]
59. Tian, L.; Qian, K.; Qi, J.; Liu, Q.; Yao, C.; Song, W.; Wang, Y. Gold nanoparticles superlattices assembly for electrochemical biosensor detection of microRNA-21. *Biosens. Bioelectron.* **2018**, *99*, 564–570. [[CrossRef](#)]
60. Su, S.; Cao, W.; Liu, W.; Lu, Z.; Zhu, D.; Chao, J.; Weng, L.; Wang, L.; Fan, C.; Wang, L. Dual-mode electrochemical analysis of microRNA-21 using gold nanoparticle-decorated MoS₂ nanosheet. *Biosens. Bioelectron.* **2017**, *94*, 552–559. [[CrossRef](#)]
61. Wang, J.; Lu, J.; Dong, S.; Zhu, N.; Gyimah, E.; Wang, K.; Li, Y.; Zhang, Z. An ultrasensitive electrochemical biosensor for detection of microRNA-21 based on redox reaction of ascorbic acid/iodine and duplex-specific nuclease assisted target recycling. *Biosens. Bioelectron.* **2019**, *130*, 81–87. [[CrossRef](#)] [[PubMed](#)]
62. Lu, J.; Wang, J.; Hu, X.; Gyimah, E.; Yakubu, S.; Wang, K.; Wu, X.; Zhang, Z. Electrochemical Biosensor Based on Tetrahedral DNA Nanostructures and G-Quadruplex-Hemin Conformation for the Ultrasensitive Detection of MicroRNA-21 in Serum. *Anal. Chem.* **2019**, *91*, 7353–7359. [[CrossRef](#)]
63. Liang, Z.; Ou, D.; Sun, D.; Tong, Y.; Luo, H.; Chen, Z. Ultrasensitive biosensor for microRNA-155 using synergistically catalytic nanoprobe coupled with improved cascade strand displacement reaction. *Biosens. Bioelectron.* **2019**, *146*, 1–8. [[CrossRef](#)] [[PubMed](#)]
64. Wu, X.; Chai, Y.; Zhang, P.; Yuan, R. An electrochemical biosensor for sensitive detection of microRNA-155: Combining target recycling with cascade catalysis for signal amplification. *ACS Appl. Mater. Interfaces* **2015**, *7*, 713–720. [[CrossRef](#)] [[PubMed](#)]
65. Wu, X.; Chai, Y.; Yuan, R.; Su, H.; Han, J. A novel label-free electrochemical microRNA biosensor using Pd nanoparticles as enhancer and linker. *Analyst* **2013**, *138*, 1060–1066. [[CrossRef](#)]
66. Azimzadeh, M.; Rahaie, M.; Nasirizadeh, N.; Ashtari, K.; Naderi-Manesh, H. An electrochemical nanobiosensor for plasma miRNA-155, based on graphene oxide and gold nanorod, for early detection of breast cancer. *Biosens. Bioelectron.* **2016**, *77*, 99–106. [[CrossRef](#)] [[PubMed](#)]